



Peter C. Iwen, Ph.D., D (ABMM), Laboratory Director
985900 Nebraska Medical Center
Omaha, NE 68198
Main: 402-559-9444
24/7 Pager: 402-888-5588

NPHL ALERT

February 5, 2024

On January 26, 2024, bioMérieux sent out a [Medical Device Recall](#) involving the BioFire® FilmArray® Gastrointestinal (GI) Panel (Part number: RFIT-ASY-0116 and RFIT-ASY-0104). These parts have a potential for increased false positive norovirus results. The required actions for this recall include:

- Distributing to all appropriate personnel
- Retaining a copy for your files
- Confirming a positive norovirus with another method
- Completing an Acknowledgement Form in Attachment A to return to bioMérieux

If you have further questions regarding this urgent message, please reach out to BioFire® technical support at biofiresupport@biomerieux.com or (800) 735-6544 and select option 5 for Product Technical Support.

Thank you,

Nebraska Public Health Lab

NPHL 24/7 Emergency Pager (402) 888-5588
Office Phone: (402) 559-9444
www.nphl.org